

# SentryCom Software



The Windows® based SentryCom Software enables you to set the values for dose and dose rate alarms and warnings, set the Cal Factor, and the data logging frequency. It also allows the user to download the accumulated exposure data, which includes time and date stamp, exposure, and power cycle identifiers for analysis. The data is stored in a convenient delimited text file for use with a spreadsheet program.

**Sentry Com 0.8.3** Settings | Calibrate | Data Log

Serial Number: 000104 | Firmware Version: 1

Retrieve Current Settings

Dose Warning Level: 0 mR | Rate Warning Level: 0 mR/h  
Dose Alarm Level: 50 mR | Rate Alarm Level: 500 mR/h  
Data Logging Frequency: 1 min

Load Defaults | Update With New Settings

Retrieve Date and Time | Update Date and Time using PC Clock  
Set Date and Time to: 8/ 1/2008 8:45:08 AM  
Today | Now

**Sentry Com 0.8.3** Settings | Calibrate | Data Log

Serial Number: 000104 | Firmware Version: 1

Retrieve Current Settings

Calibration Mode

Current Settings Are:  
Dose warning level = 0 mR  
Dose alarm level = 50 mR  
Rate alarm level = 0 mR/h  
Rate alarm level = 500 mR/h

Calibration Factor: 1.0 cps per mR/h | Dead Time: 5.00E-05 sec  
Update With New Settings

**Sentry Com 0.8.3** Settings | Calibrate | Data Log

Serial Number: 000104 | Firmware Version: 1

Retrieve Data | Complete!

Date	Time	Accumulated Exposure
05/14/10	09:18:53	0.2
05/14/10	09:28:53	0.2
05/14/10	09:38:53	0.2
05/14/10	09:48:53	0.2
05/14/10	09:58:53	0.2
	Power On	
	Clock Set	
	Power On	
	Clock Set	
	Power On	
01/01/00	00:01:19	0.0
	Power On	
	Clock Set	

Save Data...

# SE INTERNATIONAL, INC



P.O. Box 39, 436 Farm Rd. Summertown, TN 38483

1-800-293-5759 | Fax: 931-964-3564

www.seintl.com | radiationinfo@seintl.com

# RADIATION<sup>®</sup> A • L • E • R • T